

## DAP Spirometry QC Program

### *Data Submission Requirements for Each 6 Month Cycle*

Due Dates: January 15 and July 15

#### *BioQC:*

- Spirograms for each monthly BioQC test:
  - Minimum of 3 acceptable and repeatable trials
  - Numeric data plus graphs – flow-volume and volume-time for each trial
- Calibration report for each monthly BioQC result
- Completed Spirometry worksheet for BioQC 1 and/or BioQC2 (only 1 BioQC subject required/month):
  - Use largest FVC, FEV<sub>1</sub> and PEF from acceptable trials

#### *Linearity:*

- Monthly linearity results:
  - Less than 2 Liters/second
  - 4 to 6 Liters/second
  - Greater than 8 Liters/second
  - Include numerical and graphical data
- Completed Linearity worksheet

#### *Interpreted Patient Reports:*

- 10 random patient spirometrys submitted with physician interpretation:
  - Signed / dated
  - Minimum of 3 acceptable and repeatable trials
  - Numeric data plus graphs – flow-volume and volume-time for each trial
  - The above also applies to pre and post-dilator trials, if applicable
  - Calibration report or date of calibration
- Remove patient identifiers
- Number patients from 1 to 10 for your reference